



## Accessory Cable Management Mobile Stand 4800/4900 Series Cleaning Instructions

# 4800-60, 4900-60

Material number 722730  
© 2014 Welch Allyn, Inc.

Clean the Welch Allyn Accessory Cable Management Mobile Stand on a routine basis according to your facility's protocols and standards, or local regulations. Disinfect the stand according to your facility's protocols and standards, or local regulations.

The following products have been identified as being compatible with the Welch Allyn Accessory Cable Management Mobile Stand:

- Hydrogen peroxide cleaner disinfectant Wipes (Clorox Healthcare®)
- Dispatch Hospital Cleaner Disinfectant Towels with Bleach (Clorox Healthcare®)
- 70% Isopropyl alcohol
- 10% chlorine bleach



**WARNING:** Disconnect the Connex® Vital Signs Monitor from the power source before cleaning or disinfecting the stand.

### Cleaning Instructions

1. Wipe the stand using one of the cleaning agents listed above, following the cleaning agent manufacturer's guidelines.
2. Allow the stand surface to dry for a minimum of 10 minutes before using the stand.



## Accessory Cable Management Mobile Stand 4800/4900 Series Cleaning Instructions

# 4800-60, 4900-60

Material number 722730  
© 2014 Welch Allyn, Inc.

Clean the Welch Allyn Accessory Cable Management Mobile Stand on a routine basis according to your facility's protocols and standards, or local regulations. Disinfect the stand according to your facility's protocols and standards, or local regulations.

The following products have been identified as being compatible with the Welch Allyn Accessory Cable Management Mobile Stand:

- Hydrogen peroxide cleaner disinfectant Wipes (Clorox Healthcare®)
- Dispatch Hospital Cleaner Disinfectant Towels with Bleach (Clorox Healthcare®)
- 70% Isopropyl alcohol
- 10% chlorine bleach



**WARNING:** Disconnect the Connex® Vital Signs Monitor from the power source before cleaning or disinfecting the stand.

### Cleaning Instructions

1. Wipe the stand using one of the cleaning agents listed above, following the cleaning agent manufacturer's guidelines.
2. Allow the stand surface to dry for a minimum of 10 minutes before using the stand.